



Unser Zeichen  
24-P-Schu-0019

Bearbeiter  
Kerstin Schulze  
(Tel.: +493722 73 23-797)  
(E-Mail: k.schulze@slg.eu)

Datum  
13.05.2024

# Notified Body Confirmation Letter

**Reference: 24-P-Schu-0019**

**Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices**

This letter confirms that, SLG Prüf- und Zertifizierungs GmbH, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number NB 0494 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

MAG & More GmbH

Machtlfinger Straße 13

81379 München

SRN Number (if available): DE-MF-000012125

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.


In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

This confirmation shall be renewed regularly and is valid until: 2025-08-12

On behalf of the Notified Body,



2024-05-13

Ulrike Weißflog  
Certification Body Medical Devices

**Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

<b>Device name or Basic UDI-DI (under MDR application)</b>	<b>MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)</b>	<b>If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device</b>	<b>MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification</b>
<b>Stimware (1506xx) Basic UDI-DI: 426052605STIM150EH</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PowerMAG Research 30 (112269, 112270) Basic UDI-DI: 426052605PMAG1125R</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PowerMAG Clinical 30 (112249, 112250) Basic UDI-DI: 426052605PMAG1125R</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PowerMAG Clinical 100 (112259, 112260, 112261) Basic UDI-DI: 426052605PMAG1125R</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PowerMAG Research 100 (112279, 112280, 112281) Basic UDI-DI: 26052605PMAG1125R</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PowerMAG Research 100 ppTMS (112289, 112290) Basic UDI-DI: 26052605PMAG1125R</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMD70 (510519) Basic UDI-DI: 426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMR110 (510522) Basic UDI-DI: 426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMD25-decentral (510523) Basic UDI-DI: 426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMD70-pCool (510565) Basic UDI-DI: 426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494

<b>Device name or Basic UDI-DI (under MDR application)</b>	<b>MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)</b>	<b>If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device</b>	<b>MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification</b>
<b>PMD70-pCool Sham (510551)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMD45-EEG (510533)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>HANS-pCool (510570)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>HANS-pCool Sham (510571)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMD70- aCool (510575)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>PMD70- aCool Sham (510576)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>HANS-aCool (510580)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494
<b>HANS-aCool Sham (510581)</b> <b>Basic UDI-DI:</b> <b>426052605COIL5103Z</b>	Class IIa	N/A	Certificate 110143N2; NB0494

## Confirmation Letter Revision History

<b>Date</b>	<b>NB internal reference traceable to each version of the letter</b>	<b>Action</b>
2024/05/13	24-P-Schu-0019	Initial issue